



# Confirmation

**Braincon GesmbH & CO KG**

**Grinzinger Allee 5**

**1190 Wien**

**Austria**

**DCXF (produced by SANOSIL) in combination with DCXplus**

was tested according to

**EN 17272:2020**

Methods of airborne room disinfection by automated process - Determination of bactericidal, mycobactericidal, sporicidal, fungicidal, yeasticidal, virucidal, tuberculocidal and phagocidal activities.

For the product, a claim according to EN 17272:2020 of **bactericidal, mycobactericidal, sporicidal, fungicidal and virucidal** for the **human medicine** and the **areas food, industry, household and public facilities sectors** were fulfilled.

This is confirmed by the test report number: B 30876 - B 30876f.

Identification Number: 2330876/075

Bischofshofen, 2023-10-19

A handwritten signature in blue ink, appearing to read "Werner", is written over a horizontal line.

Prof. Dr. med. H.-P. Werner  
Medical Specialist for Hygiene & Microbiology